

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Abbott Rapid Diagnostics	Panbio™ COVID-19 Ag Rapid Test	1232	<i>Prospective clinical field studies</i>					
			<p>BE: Small-scale head-to-head comparison of 5 RATs in Belgian hospital lab. Panbio overall sensitivity (Ct range 14,6 – 35,5): 45/57 samples (79%). Sensitivity for Ct≤25: 17/18 samples. Overall specificity 100%.</p> <p>NL: 1367 and 208 subjects were enrolled in Utrecht and Aruba, respectively. Specificity of the Panbio™ COVID-19 Ag Rapid Test was 100% (95%CI: 99.7–100%) in both settings. Test sensitivity was 72.6% (95%CI: 64.5–79.9%) in the Netherlands and 81.0% (95% CI: 69.0–89.8%) in Aruba. Restricting RT-qPCR test positivity to Ct-values <32 yielded test sensitivities of 95.2% (95%CI: 89.3–98.5%) in Utrecht and 98.0% (95%CI: 89.2–99.95%) in Aruba.</p> <p>PT: 83 samples from symptomatic individuals (27 PCR positive and 56 negative by PCR) were tested. Sensitivity 63% (95%IC 42-81); specificity 100% (95%IC 94-100). LoD TCID50/ml 1,38 x 10² and CT<24.</p> <p>SE: Karolinska hospital evaluation of Lot 41ADF061A. Patient samples: 95 PCR positive, 150 negative. No detailed sample description available. Sensitivity 59%, specificity 100%. Sensitivity Ct<25 = 90.2%.</p> <p>FIND evaluation studies DE (10 Dec 2020) 1108 samples, NP swab. Clinical sensitivities: Days ≤ 7: 90.8%; Ct ≤ 33: 88.3%; Ct < 25: 95.8%. Clinical specificity: 99.9% CH (10 Dec 2020) 535 samples, NP swab. Clinical sensitivities: Days < 7: 85.6%; Ct ≤ 33: 89.7%; Ct < 25: 96.8%. Clinical specificity: 100% India (25 June 2021) 526 samples, NP swab. Clinical sensitivities: Days < 7: 61.3%-100%; Ct ≤ 33: 74.2%-86.7%; Ct < 25: 91.9%-100%. Clinical specificity: 100%</p>	<p>91.4% sensitivity 99.8% specificity NP swab (Ct ≤ 33)</p> <p>98.1% sensitivity 99.8% specificity Nasal swab (Ct ≤ 33)</p>	<p>BE, DE^[2], ES, FI, NL^[5], PT, SE</p> <p>CH, India, NO, UK</p>	Nucleo-protein	Nasal swab, Nasopharyngeal swab	17 February 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
			<i>Retrospective in vitro studies</i>					
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity of 99.8%					
ABIOTEQ	Cora Gentest-19	2374	<i>Retrospective in vitro study</i>	Sensitivity 98,7%, Specificity 99,8%	DE ^[2]	Nucleo-capsid protein	Anterior nasal swab, Nasal swab, Nasopharyngeal swab, Oropharyngeal swab, Throat swab	20 October 2021
			<i>Retrospective in vitro study</i>					
AccuBioTech Co.,Ltd	Accu-Tell SARS-CoV-2 Ag Cassette	2579	DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.2%	Sensitivity: 95.7% Specificity: 99.2%	DE ^[2]	Nucleo-capsid protein	Nasopharyngeal swab	20 October 2021
			<i>Prospective clinical field study</i>					
Acon Biotech (Hangzhou) Co., Ltd	Flowflex SARS-CoV-2 Antigen Rapid Test	1457	FIND evaluation CH (9 June 2021) 279 samples, nasal swab. Clinical sensitivities: Days ≤ 7: 92.2%; Ct ≤ 33: 98.3%; Ct ≤ 25: 100%. Clinical specificity: 99.5%	Nasal swab Clinical Sensitivity: 97.1 % Clinical Specificity: 99.5 %	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab	14 July 2021
			<i>Retrospective in vitro study</i>	NP swab Clinical Sensitivity: 97.6 % Clinical Specificity: 99.4 %	CH, UK			
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 99.54%					
ACON Laboratories, Inc.	Flowflex SARS-CoV-2 Antigen Rapid Test	1468	<i>Retrospective in vitro study</i>	96.9% sensitivity 98.7% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab	10 May 2021
			<i>Retrospective in vitro study</i>					
AESKU.DIAGNOSTICS GmbH & Co, KG	AESKU.RAPID SARS-CoV-2	2108	DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 84% at Ct ≤ 25; Manufacturer specificity: 98%	96% sensitivity 98% specificity NP swab	DE ^[2]	Nucleo-protein	Nasal swab, Throat swab	10 May 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Affimedix Inc.	TestNOW® - COVID-19 Antigen Test	2130	<i>Retrospective in vitro study</i>	95% sensitivity 99.2% specificity NP swab 98.1% sensitivity 100% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99,2%					
AMEDA Labordiagnostik GmbH	AMP Rapid Test SARS-CoV-2 Ag	1304	<i>Retrospective in vitro study</i>	97.3% sensitivity NP swab 97.3% sensitivity Nasal swab 100% specificity	DE ^[2] CH, UK	Nucleo-protein	Nasal swab, Nasopharyngeal swab	17 February 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%					
Anbio (Xiamen) Biotechnology Co., Ltd	Rapid COVID-19 Antigen-Test (colloidal Gold)	1822	<i>Retrospective in vitro study</i>	99.27% sensitivity, 100% specificity Nasal swab	DE ^[2]	Unknown	Nasal swab, Throat swab	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%					
Anhui Deep Blue Medical Technology Co., Ltd	COVID-19 (SARS-CoV-2) Antigen Test Kit (Colloidal Gold)	1736	<i>Retrospective in vitro study</i>	Nasal/OP swab: 96,4% sensitivity, 99,8% specificity NP swab: 95,7% sensitivity, 99,3% specificity	DE ^[2] UK	Nucleo-protein	Nasal swab, ! Other	10 May 2021
	COVID-19 (SARS-CoV-2) Antigen Test Kit (Colloidal Gold) – Nasal swab	1815	<i>Retrospective in vitro study</i>	96.4 % sensitivity 99.8 % specificity Nasal swab	DE ^[2] UK	Nucleo-protein	Anterior nasal swab, Nasal swab	10 May 2021
Anhui Formaster Biosci Co., Ltd.	New Coronavirus (COVID-19) Antigen Rapid Test	2089	<i>Retrospective in vitro study</i>	sensitivity: 95.15%, specificity: 98.5%	DE ^[2]	Nucleo-protein	Nasopharyngeal swab, Oropharyngeal swab	20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.5%					
ArcDia International Ltd	mariPOC SARS-CoV-2	768	<i>Prospective clinical field study</i>	100% sensitivity 100% specificity Nasopharyngeal swab	FI	Nucleo-protein	Nasopharyngeal swab	10 May 2021
			FI : Clinical performance of the test was evaluated against qRT-PCR with nasopharyngeal swab specimens collected from patients suspected of acute SARS-CoV-2 infection. Sensitivity of the mariPOC test was 100.0% (13/13) directly from swab specimens and 84.4% (38/45) from swab specimens in undefined transport mediums. Specificity of the test was 100.0% (201/201).					

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS- CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
ArcDia International Oy Ltd	mariPOC Respi+	2078	<i>Retrospective in vitro study</i>	100 % sensitivity 100 % specificity NP swab	FI	Nucleo- protein	Nasopharyngeal swab	14 July 2021
			FI: Validated in several laboratories (studies not published), meeting criteria.					
ArcDia International Oy Ltd	mariPOC Quick Flu+	2079	<i>Retrospective in vitro study</i>	100 % sensitivity 100 % specificity NP swab	FI	Nucleo- protein	Nasopharyngeal swab	14 July 2021
			FI: Validated in several laboratories (studies not published), meeting criteria.					
Artron Laboratories Inc.	Artron COVID-19 Antigen Test	1618	<i>Retrospective in vitro study</i>	96.67% sensitivity, Nasal swab 91.67% sensitivity, NP swab 100 % specificity Nasal/NP swab	DE ^[2]	Nucleo- protein	Nasal swab, Nasopharyngeal swab	14 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%					
Asan Pharmaceutical Co., Ltd	Asan Easy Test COVID-19 Ag	1654	<i>Retrospective in vitro study</i>	94.67% sensitivity, 97.71% specificity Nasal swab	DE ^[2]	Unknown	Nasal swab	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 97.71%					
Assure Tech. (Hangzhou) Co., Ltd.	ECOTEST COVID-19 Antigen Rapid Test Device	770	<i>Retrospective in vitro study</i>	92.5 % sensitivity 99.2 % specificity Nasal/NP/OP swab	DE ^[2] UK	Nucleo- protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	14 July 2021
	ECOTEST COVID-19 Antigen Rapid Test Device	2350	<i>Retrospective in vitro study</i>					
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 95% at Ct ≤ 25; Manufacturer specificity: 99.2%					
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 95% at Ct ≤ 25; Manufacturer specificity: 99.1%	Sensitivity: 97.7%, Specificity: 99.1% NP and OP swab	DE ^[2] UK	Nucleo- protein	Nasopharyngeal swab, Oropharyngeal swab	23 July 2021
Atlas Link Technology Co. Ltd.	NOVA Test® SARS-CoV-2 Antigen Rapid Test Kit (Colloidal Gold Immunochromatography)	2010	DE: 97.6% sensitivity, 99.2% specificity	98.5 % sensitivity 99.4 % specificity Nasal/OP swab	DE ^[2] CH	Nucleo- protein	Nasal swab, Oropharyngeal swab	10 May 2021 ¹⁸

¹⁸ This rapid antigen test, device ID 2010, was removed from the EU common list on 8 December 2021. The grace period will end on 2 Feb 2022, 23:59 CET.

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Avalun	Ksmart® SARS-COV2 Antigen Rapid Test	1800	<i>Retrospective in vitro study</i>	Sensitivity: 93.18% Specificity: 99.32% NP swab	DE ^[2]	Unknown	Nasopharyngeal swab	7 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 99,32%					
AXIOM Gesellschaft für Diagnostica und Biochemica mbH	COVID-19 Antigen Rapid Test	2101	<i>Retrospective in vitro study</i>	98% sensitivity 100% specificity NP/Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Throat swab	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%					
Azure Biotech, Inc.	COVID-19 Antigen Rapid Test Device	1906	<i>Retrospective in vitro study</i>	95% sensitivity 99.2% specificity NP swab	DE ^[2]	Unknown	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (NP swab): Sensitivity of 86% at Ct ≤ 25; Manufacturer specificity: 99.2%					
Becton Dickinson	BD Veritor™ System for Rapid Detection of SARS CoV 2	1065	<i>Prospective clinical field studies</i>	Clinical Sensitivity: 91.1 % Clinical Specificity: 99.6 % Nasal swab	DE ^[2] , ES, NL, SE	Nucleo-protein	Nasal swab	7 July 2021
			ES: Prospective study in four Spanish hospitals (n = 476); 108 positive samples, 368 negative samples. Sensitivity: 92%, specificity: 98.6%.					
			NL: Independent field study in symptomatic individuals (n=979, PCR positive n=161) - sampling was Nasal mid-turbinate + OP swab. Sensitivity overall: 79.5% - Sensitivity Ct<30: 93.2% - Specificity overall: 99.8%					
			SE: Karolinska hospital evaluation of Lot 0255648. Patient samples: 95 PCR positive, 150 negative. No detailed sample description available. Sensitivity 45%, specificity 97%. Sensitivity Ct<25 = 87.8%.					
			<i>Retrospective in vitro study</i>					
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 83% at Ct ≤ 25; Manufacturer specificity: 99.6%					

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Becton Dickinson	BD Kit for Rapid Detection of SARS-CoV-2	2282	<i>Prospective clinical field studies</i> ES: Prospective study in four Spanish hospitals (n = 476); 108 positive samples, 368 negative samples. Sensitivity: 92%, specificity: 98.6%. NL: Independent field study in symptomatic individuals (n=979, PCR positive n=161) - sampling was Nasal mid-turbinate + OP swab. Sensitivity overall: 79.5% - Sensitivity Ct<30: 93.2% - Specificity overall: 99.8%	Clinical Sensitivity: 91.1 % Clinical Specificity: 99.6 % Nasal swab	ES, NL	Nucleo-protein	Nasal swab	10 November 2021
Beijing Hotgen Biotech Co., Ltd	Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold)	1870	<i>Prospective clinical field study</i> FIND evaluation Brazil (15 September 2021) 453 samples, nasal swab. Clinical sensitivities: Days ≤ 7 : 90.1%; Ct ≤ 33 : 89.5%; Ct ≤ 25 : 95.5%. Clinical specificity: 100% UK (15 September 2021) 248 samples, NP swab. Clinical sensitivities: Days ≤ 7 : 84.4%; Ct ≤ 33 : 80.6%; Ct ≤ 25 : 82.8%. Clinical specificity: 99.4% <i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25 ; Manufacturer specificity: 99.76%	97.1% sensitivity 99.76% specificity	DE ^[2]	Nucleo-protein	Nasal swabs, Throat swabs, ! Saliva	10 May 2021
Beijing Jinwofu Bioengineering Technology Co., Ltd.	Novel Coronavirus (SARS-CoV-2) Antigen Rapid Test Kit	2072	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25 + Manufacturer specificity: 100%	96.88 % sensitivity 100 % specificity Nasal/ NP/ OP swab	DE ^[2]	Nucleo-protein	Anterior nasal swab, Nasal swab, Nasopharyngeal swab, Oropharyngeal swab ! Saliva	14 July 2021
Beijing Lepu Medical Technology Co., Ltd	SARS-CoV-2 Antigen Rapid Test Kit	1331	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25 ; Manufacturer specificity: 99.26%	92.00% sensitivity, 99.26% specificity Nasal swab	DE ^[2]	Unknown	Nasal swab, Nasopharyngeal swab	17 February 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Beijing O&D Biotech Co., Ltd.	COVID-19 Antigen Rapid Test	2494	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.67%	sensitivity: 92.17%, specificity: 98.67 %	DE ^[2]	Nucleo-capsid protein	Nasal swab, Oropharyngeal swab	20 October 2021
Beijing Wantai Biological Pharmacy Enterprise Co., Ltd	Wantai SARS-CoV-2 Ag Rapid Test (FIA)	1484	No data available	96.6% sensitivity, Nasal swab 93 % sensitivity 97% specificity NP, OP swabs		Nucleo-protein	Nasal swab ! Other	17 February 2021 ¹⁹
Beijing Wantai Biological Pharmacy Enterprise Co., Ltd	Wantai SARS-CoV-2 Ag Rapid Test (colloidal gold)	1485	<i>Prospective clinical field study</i> CZ: Independent prospective study by Public Health Institute Ostrava (CZ), including NP swabs from unselected symptomatic and asymptomatic participants. Sensitivity 80.6%, specificity 98.5% on 155 pos. and 325 neg. samples (as resulting by RT-PCR). Ct not reported. N total = 480 <i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.2%	93.2% sensitivity 98.2% specificity Nasal swab	CZ, DE ^[2]	Unknown	Anterior nasal swab, Nasal swab, Nasopharyngeal swab, Oropharyngeal swab ! Saliva	14 July 2021
BioGnost Ltd	CoviGnost AG Test Device 1x20	2247	<i>Retrospective in vitro study</i> HR: 300 NP samples (retrospective), symptomatic (<7 dps): 200 PCR+ samples (range Ct 16-30), Ct<30: sensitivity 96.5% 100 PCR- samples: specificity 100%	Sensitivity: 96%, Specificity: 99% NP swab	HR	Unknown	Nasopharyngeal swab	23 July 2021
BIOHIT HealthCcare (Hefei) Co., Ltd.	SARS-CoV-2 Antigen Rapid Test Kit (Fluorescence Immunochromato-graphy)	1286	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.9%	Sensitivity: 96.77% Specificity: 98.9% NP/OP swab	DE ^[2]	Nucleo-capsid protein	Anterior nasal swab	23 July 2021

¹⁹ This rapid antigen test, device ID 1484, was removed from the EU common list on 20 October 2021. The grace period will end on 15 Dec 2021, 23:59 CET

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
BIOHIT HealthCcare (Hefei) Co., Ltd.	SARS-CoV-2 Antigen Rapid Test (Colloidal Gold Method)	2230	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.49%	Sensitivity: 96.12%, Specificity: 99.49 %	DE ^[2]	Nucleo-capsid protein	Nasal swab	8 December 2021
BioMaxima SA	SARS-CoV-2 Ag Rapid Test	2035	<i>Prospective clinical field studies</i> FR: NP swabs, Diagnostic sensitivity: 96,4% (80/83) (95% CI: 89,8-99,2%); diagnostic specificity: 99,2%, (120/121) PL: Evaluation of the test was performed on 480 samples of nasopharyngeal swabs taken from patients with symptoms of COVID-19 and from people in contact with an infected person but without symptoms of infection. Positive results of the antigen test were obtained in 205 patients and in the molecular test 213 people. On the other hand, negative results of the antigen test were obtained in 275 people and in the molecular test 267 people. The above results permitted calculation of diagnostic sensitivity, which was 93.43% (95% CI: 91.61%~97.19%) and diagnostic specificity, which was 97.75% (95% CI: 93.74%~98.92%) <i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99%	Sensitivity: 95% Specificity: 99% NP Swab	DE ^[2] , FR, PL	Nucleo-protein	Nasopharyngeal swab	23 July 2021
Biomerica Inc.	Biomerica COVID-19 Antigen Rapid Test (nasopharyngeal swab)	1599	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99,7%	Clinical Sensitivity: 94.7%; Clinical specificity: 99.7% Nasal/NP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab	7 July 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
BIONOTE	NowCheck COVID-19 Ag Test	1242	<i>Prospective clinical field study</i>	Clinical Sensitivity: 90.91 % Clinical Specificity: 99.43 % Nasal swab, NP swab	DE ^[2] Brazil	Unknown	Nasal swab, Nasopharyngeal swab	7 July 2021
			FIND evaluation Brazil (20 April 2021) 400 samples, NP swab. Clinical sensitivities: Days \leq 7: 92.2%; Ct \leq 33: 91.4%; Ct \leq 25: 94.8%. Clinical specificity: 97.3%					
			Brazil (30 March 2021) 218 samples, Nasal/NP swab. Clinical sensitivities: Days \leq 7: 92.5% (N/NP); Ct \leq 33: 97.2% (N/NP); Ct $<$ 25: 100% (N/NP); Clinical specificity: 98.6%					
			<i>Retrospective in vitro study</i>					
BIO-RAD	CORONAVIRUS AG RAPID TEST CASSETTE	2031	DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct \leq 25; Manufacturer specificity: 98,6%	Clinical Sensitivity: 98% (NP: 98,32% / Nasal: 97,25%) Clinical Specificity: 99% (NP: 99,6% / Nasal: 100%)	ES	Nucleo-protein	Nasal swab, Nasopharyngeal swab	7 July 2021
			<i>Prospective clinical field studies</i> ES⁷: <ul style="list-style-type: none"> Prospective study; 96 positive samples and 269 negative samples. Sensitivity 94%. Specificity 99.2%. No Ct distribution specified. NP swab: sensitivity 98,3%; specificity 99,6% (119 positive samples, 746 negative samples) Nasal swab: sensitivity 97,2%; specificity 100% (109 positive samples, 128 negative samples) 					

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
BIOSYNEX SWISS S.A.	BIOSYNEX COVID-19 Ag BSS	1223	<i>Prospective clinical field studies</i>	96% sensitivity, 100% specificity, NP/Nasal swab	BE, DE ^[2] , FR, NL ^[5] , SE CH	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab	17 February 2021
			BE^[6]: Small-scale head-to-head comparison of 5 RATs in Belgian hospital lab. Biosynex overall sensitivity (Ct range 14.6 – 35.5): 52/58 samples (89.7%). Sensitivity for Ct≤25: 18/18 samples. Overall specificity only 46.2%, but this is probably linked to the use of transport medium instead of the swab included in the kit.					
			FR: NP swabs, prospective study (71/71) : sensitivity 100% (45/45, specificity 100%					
			NL: Independent field study, mainly symptomatic individuals (n=568, PCR positive n=39), NP swab; sensitivity Ct≤30: 96.0%, sensitivity ≤25: 100%; specificity overall: 100%					
			NL: Independent field study, symptomatic individuals (n=270, PCR positive n=17), NP+OP swab; sensitivity Ct≤30: 94.1%, sensitivity Ct≤25: 100%; specificity overall: 100%					
			SE: Karolinska hospital evaluation of Lot 20100103. Patient samples; 95 PCR positive, 150 negative. No detailed sample description available. Sensitivity 76%, specificity 96%. Sensitivity Ct<25 = 100%.					
			<i>Retrospective in vitro study</i>					
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%					
BIOSYNEX SA	BIOSYNEX COVID-19 Ag+ BSS	1494	<i>Prospective clinical field study</i>					
			FR: Validation study data: 125 positive and 118 negative samples; sensitivity 96%, specificity: 99%					
				Clinical Sensitivity: 97.5 % Specificity: 99% Nasal swab, NP swab	FR UK	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab	7 July 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
BIOTEKE CORPORATION (WUXI) CO., LTD	SARS-CoV-2 Antigen Test Kit (colloidal gold method)	2067	<i>Retrospective in vitro study</i>	96.49 % sensitivity 99.28 % specificity OP/NP swab	DE ^[2]	Nucleo-protein	Nasopharyngeal swab, Oropharyngeal swab	14 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 95% at Ct ≤ 25; Manufacturer specificity: 99.28%					
Biotical Health S.L.U. BIOTICAL HEALTH S.L.U	biotical SARS-CoV-2 Ag Card	2013	<i>Retrospective in vitro study</i>	Sensitivity: 96%, Specificity: 99% NP swab	BE	Nucleo-protein	Nasopharyngeal swab	23 July 2021
			BE: Validation study 1: sensitivity 91.7% for Ct<25; Validation study 2: 94% for Ct<25. Manufacturer specificity: 99%					
Boditech Med Inc	AFIAS COVID-19 Ag	1989	<i>Prospective clinical field study</i>	Sensitivity: 91.7%, Specificity: 98.7% NP swab	NL	Nucleo-protein	Nasopharyngeal swab	23 July 2021
			NL: Independent field study in mild symptomatic (n= 427, PCR positive: 106); unknown swab, overall sensitivity: 81.1%, sensitivity Ct <30: 96.4%; specificity: 100%,					
BTNX Inc	Rapid Response COVID-19 Antigen Rapid Test	1236	<i>Retrospective in vitro study</i>	90.2% sensitivity 100% specificity NP swab, NP swab, OP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (NP swab): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%					
CerTest Biotec	CerTest SARS-CoV-2 Card test	1173	<i>Prospective clinical field study</i>	92.9% sensitivity 99.6% specificity NP swab	DE ^[2] , ES	Unknown	Nasopharyngeal swab	17 February 2021
			ES: Ct ≤ 25, sensitivity: 94,0%; sensitivity for samples within the first 5 days after symptom onset: 84,8%; 150 positive samples, 170 negative samples					
Chil Tibbi Malzeme Sanayi ve Ticaret Limited Şirketi	CHIL COVID-19 Antigen Rapid Test (Nasopharyngeal / Oropharyngeal Swab-Casette)	1691	<i>Retrospective in vitro study</i>	Sensitivity 99.01% Specificity: 99.57%	DE ^[2]	Nucleo-protein	Nasopharyngeal swab, Oropharyngeal swab	20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.57%					
Chongqing M&D Biotechnology Co. Ltd	2019-nCoV Antigen Test Kit	2150	<i>Retrospective in vitro study</i>	sensitivity: 91.53%, specificity:100%	DE ^[2]	Nucleo-protein	Nasopharyngeal swab	20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 95% at Ct ≤ 25; Manufacturer specificity: 100%					

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Core Technology Co., Ltd	Coretests COVID-19 Ag Test	1919	<i>Retrospective in vitro study</i>	98.1% sensitivity 99.6% specificity NP swab	DE ^[2]	Nucleo-protein	Nasopharyngeal swab	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 88% at Ct ≤ 25; Manufacturer specificity: 99.6%					
CTK Biotech, Inc	OnSite COVID-19 Ag Rapid Test	1581	<i>Prospective clinical field study</i>	Clinical Sensitivity: 92.3 % Clinical Specificity: 100 % Nasal, NP swab	DK, ES	Nucleo-protein	Nasal swab, Nasopharyngeal swab	7 July 2021
			DK: 107 samples; Nasal swab - clinical sensitivity 86%; (from asymptomatic and mild symptomatic individuals), Clinical specificity: 100%					
DDS DIAGNOSTIC	Test Rapid Covid-19 Antigen (tampon nazofaringian)	1225	<i>Prospective clinical field study</i>	98.77% sensitivity 99.03% specificity Nasal swab	RO China	Unknown	Nasal swab	10 May 2021
			RO: Clinical study based on 228 COVID-19 positive samples and 597 COVID-19 negative samples. All the samples were confirmed using PCR (Applied Biosystems™ 7500 and SLAN®- 96P) and clinical symptoms. The relative sensitivity of Rapid Test COVID-19 Antigen (Nasopharyngeal Swab) was 99.56%, the relative specificity was 99.66%, and the accuracy was 99.64% compared to the qRT-PCR result.					
DIALAB GmbH	DIAQUICK COVID -19 Ag Cassette	1375	BE: Z20401CE: 93.2% sensitivity, 100% specificity, NP swab Z20601CE: 96.4% sensitivity, 99.2% specificity, NP swab	NP swab	BE	Unknown	Nasopharyngeal swab	10 May 2021 ²⁰
DNA Diagnostic	COVID-19 Antigen Detection Kit	2242	<i>Retrospective in vitro study</i>	Sensitivity: 93.8%, Specificity: 99.6% Nasal swab	DE ^[2] UK	Nucleo-protein	Nasal swab	23 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct < 25; Manufacturer specificity: 99.56%					

²⁰ This rapid antigen test, device ID 1375, was removed from the EU common list on 8 December 2021. The grace period will end on 2 Feb 2022, 23:59 CET.

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Dräger Safety AG & Co. KGaA	Dräger Antigen Test SARS-CoV-2	2273	<i>Prospective clinical field studies</i>	Sensitivity: 96.1% (Ct values ≤25) Specificity: 99.6%	DE ^[2] CH	Nucleo-capsid protein	Nasal swab	20 October 2021
			DE: Independent prospective study, mainly symptomatic <7 dps (n=378, PCR positive = 70), self-collected nasal swab; sensitivity overall: 88.6%, sensitivity Ct<26: 96.8%; specificity overall: 99.7%					
			CH: Independent prospective study, mainly symptomatic ≤7 dps (n=464, PCR positive = 57), self-collected nasal swab; sensitivity Ct<30: 85.1%, sensitivity Ct<26: 90.0%; specificity overall: 100%					
			<i>Retrospective in vitro study</i>					
Dynamiker Biotechnology(Tianjin) Co., Ltd.	Dynamiker SARS-CoV-2 Ag Rapid Test	2533	<i>Retrospective in vitro study</i>	sensitivity: 95.7%, specificity: 99.1%	DE ^[2]	Nucleo-capsid protein	Nasopharyngeal swab, Oropharyngeal swab	20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer Specificity: 99.1%					
Edinburgh Genetics Limited	Edinburgh Genetics ActivXpress+ COVID-19 Antigen Complete Testing Kit	1243	<i>Prospective clinical field study</i>	Clinical Sensitivity 97.27% NP swab Clinical Specificity 99.62% NP swab	DE ^[2] Peru	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	14 July 2021
			FIND evaluation Peru (26 April 2021) 120 samples, NP swab. Clinical sensitivities: Days ≤ 7: 62%; Ct ≤ 33: 75%; Ct ≤ 25: 100%. Clinical specificity: 100%					
			<i>Retrospective in vitro study</i>					
Eurobio Scientific	EBS SARS-CoV-2 Ag Rapid Test	1739	<i>Prospective clinical field study</i>	Clinical Sensitivity: 95.7 % Nasal swab	DE ^[2] , FR	Nucleo-protein	Nasal swab	7 July 2021
			FR: Validation study data: 119 positive and 125 negative samples; sensitivity 93%, specificity: 99%					

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Genrui Biotech Inc	SARS-CoV-2 Antigen Test Kit (Colloidal Gold)	2012	<i>Retrospective in vitro study</i>	Sensitivity: 91.15% Specificity: 99.02% Nasal/NP/OP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	7 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 99,02%					
GenSure Biotech Inc	GenSure COVID-19 Antigen Rapid Test Kit	1253	<i>Retrospective in vitro study</i>	96.86% sensitivity 100% specificity Nasal swab	DE ^[2] UK	Unknown	Nasal swab	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 100%					
Getein Biotech, Inc	SARS-CoV-2 Antigen (Colloidal Gold)	1820	<i>Retrospective in vitro study</i>	97.06% sensitivity 98.71% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab ! Saliva	14 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.71%					
Getein Biotech, Inc.	One Step Test for SARS-CoV-2 Antigen (Colloidal Gold)	2183	<i>Retrospective in vitro study</i>	97.06% sensitivity 98.71% specificity Nasal swab	DE ^[2] UK	Nucleo-protein	Nasal swab ! Saliva	16 June 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 90% at Ct ≤ 30 and 100% at Ct ≤ 25; Manufacturer specificity: 98.71%					
Goldsite Diagnostic Inc.	SARS-CoV-2 Antigen Kit (Colloidal Gold)	1197	<i>Retrospective in vitro study</i>	93.04% sensitivity; 100% specificity Nasal swab	FR, DE ^[2] , ES UK	Unknown	Nasal swab ! Other	14 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%					
Green Cross Medical Science Corp.	GENEDIA W COVID-19 Ag	1144	<i>Retrospective in vitro study</i>	100% sensitivity 90.1% sensitivity 100% specificity NP swab, Anterior nasal swab	DE ^[2]	Nucleo-protein	Anterior nasal swab, Nasopharyngeal swab	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 83% at Ct ≤ 25; Manufacturer specificity: 100%					
Guangdong Hecin Scientific, Inc.	2019-nCoV Antigen Test Kit (colloidal gold method)	1747	<i>Retrospective in vitro study</i>	96.23% sensitivity 99.07% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasopharyngeal swab	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 82% at Ct ≤ 25; Manufacturer specificity: 99.07%					

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Guangdong Longsee Biomedical Co., Ltd.	2019-nCoV Ag Rapid Detection Kit(Immuno-Chromatography)	1216	<i>Retrospective in vitro study</i>	99.72% sensitivity 99.5% specificity NP/OP swab	DE ^[2]	Unknown	Nasopharyngeal swab, Oropharyngeal swab	14 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.5%					
Guangdong Wesail Biotech Co. Ltd	COVID-19 Ag Test Kit	1360	<i>Retrospective in vitro study</i>	90% sensitivity 98% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab	17 February 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98%					
Guangzhou Decheng Biotechnology CO., Ltd	V-CHEK, 2019-nCoV Ag Rapid Test Kit (Immuno-chromatography)	1324	<i>Retrospective in vitro study</i>	Clinical Sensitivity: 95.83% Specificity 99.57% Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab	7 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 99,5%					
Guangzhou Wondfo Biotech Co., Ltd	Wondfo 2019-nCoV Antigen Test (Lateral Flow Method)	1437	<i>Prospective clinical field study</i>	Sensitivity: 87.12% Specificity: 99.74%	DE ^[2] CH, UK	Unknown	Nasopharyngeal swab	10 May 2021
			FIND evaluation CH (25 Feb 2020) 328 samples, NP swab. Clinical sensitivities: Days ≤ 7: 85.7%; Ct ≤ 33: 92.2%; Ct ≤ 25: 100%. Clinical specificity: 100%					
			Brazil (10 Oct 2021) 237 samples, NP swab. Clinical sensitivities: Days ≤ 7: 90.4%; Ct ≤ 33: 89.3%; Ct ≤ 25: 96.7%. Clinical specificity: 98.8%					
			<i>Retrospective in vitro study</i>					
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 88% at Ct ≤ 25; Manufacturer specificity: 99.74%					
			<i>Prospective clinical field study</i>	93,40% sensitivity, 99,90% specificity NP swab	FR	Unknown	Nasopharyngeal swab	10 May 2021
Hangzhou AllTest Biotech Co., Ltd	COVID-19 Antigen Rapid Test	1257	FR: Prospective study, sensitivity 96,4% (80/83), specificity 99,2% (120/121)					
			<i>Retrospective in vitro study</i>	Sensitivity: 93.2%, Specificity: 99.2% Nasal swab	DE ^[2]	Nucleo-capsid protein	Nasal swab	8 December 2021
Hangzhou Biotest Biotech Co., Ltd	COVID-19 Antigen Rapid Test Cassette(Nasal Swab)	1876	DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.2%					

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS- CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Hangzhou Clongene Biotech Co., Ltd.	COVID-19 Antigen Rapid Test Casette	1610	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 94,4% at Ct ≤ 25; Manufacturer specificity: 100%	Clinical Sensitivity: 91.4 % Clinical Specificity: 100 % NP swab	DE ^[2] UK	Nucleo- protein	Nasopharyngeal swab	7 July 2021
	Covid-19 Antigen Rapid Test Kit	1363	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 94,4% at Ct ≤ 25; Manufacturer specificity: 100%	98.5% (Ct<33) sensitivity Nasal swab	DE ^[2] CH	Nucleo- protein	Nasal swab	17 February 2021
	COVID-19/Influenza A+B Antigen Combo Rapid Test	1365	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 94,4% at Ct ≤ 25; Manufacturer specificity: 100%	91% sensitivity 100% specificity NP swab	DE ^[2]	Nucleo- protein	Nasopharyngeal swab	10 May 2021
Hangzhou Immuno Biotech Co., Ltd	Immunobio SARS-CoV-2 Antigen ANTERIOR NASAL Rapid Test Kit (minimal invasive)	1844	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 88% at Ct ≤ 25; Manufacturer specificity: 100%	94% sensitivity 100% specificity Nasal swab, NP swab	DE ^[2]	Nucleo- protein	Nasal swab, Nasopharyngeal swab	10 May 2021
	SARS-CoV2 Antigen Rapid Test	2317	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 88% at Ct ≤ 25; Manufacturer specificity: 100%	Clinical Sensitivity: 98 % Clinical Specificity: 100 % Anterior nasal swab, NP swab, OP swab,	DE ^[2]	Nucleo- protein	Anterior nasal swab, Nasopharyngeal swab, Oropharyngeal swab ! Sputum	10 May 2021
Sigmed Sp. z o.o.	Redtest Professional Sars-CoV-2 Antigen Rapid Test (Covid-19 Ag)	2256	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 88% at Ct ≤ 25; Manufacturer specificity: 100%	sensitivity: 98,13%, specificity: 100%	DE ^[2]	Nucleo- capsid protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	8 December 2021
Hangzhou Laihe Biotech Co.	LYHER Novel Coronavirus (COVID-19) Antigen Test Kit (Colloidal Gold)	1215	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 99,7%	Clinical Sensitivity: 95.07% % Clinical Specificity: 99.74% Nasal swab	DE ^[2] UK	Unknown	Nasal swab	10 May 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Hangzhou Lysun Biotechnology Co. Ltd	COVID-19 Antigen Rapid Test Device (Colloidal Gold)	2139	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%	96.46% sensitivity 100% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab	10 May 2021
Hangzhou Sejoy Electronics & Instruments Co.Ltd	SARS-CoV-2 Antigen Rapid Test Cassette	1945	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%	Sensitivity: 94.5%, Specificity:100% Nasal swab	DE ^[2]	Nucleo-capsid protein	Nasal swab	8 December 2021
Hangzhou Testsea Biotechnology Co., Ltd.	Covid-19 Antigen Test Cassette	1392	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (NP swab): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.4%	92.1% sensitivity 98.1% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab	10 May 2021
Healgen Scientific	Coronavirus Ag Rapid Test Cassette	1767	<i>Prospective clinical field studies</i> NL: Independent clinical field study, symptomatic individuals (n=417, PCR positive n=70), NP swab; sensitivity overall: 75.7%, sensitivity Ct≤30: 85.2%, sensitivity Ct≤25: 90.7%; specificity overall: 100% NL: Independent clinical field study, symptomatic individuals (n=240, PCR positive n=21), NP+OP swab; sensitivity overall: 85.7%, sensitivity Ct≤30: 89.5%, sensitivity Ct≤25: 100%; specificity overall: 100% NL: Independent clinical field study, symptomatic individuals (n=94, PCR positive n=18), NP+OP swab in VTM; sensitivity overall: 90.0%, sensitivity Ct≤30: 100%, sensitivity Ct≤25: 100%; specificity overall: 97.3% <i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut (NP swab): Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%	98.32 % sensitivity 99.6% specificity (NP swab) 97.25% sensitivity 100% specificity (Nasal swab)	DE ^[2] , NL ^[5]	Nucleo-proteins, S1, S1-RBD, S2	Nasal swab, Nasopharyngeal swab	17 February 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Siemens Healthineers	CLINITEST Rapid COVID-19 Antigen Test	1218	<i>Prospective clinical field studies</i>	98.32% sensitivity (NP swab) 97.25% sensitivity (Nasal swab) 100% specificity	DE ^[2] , ES, NL ^[5]	Nucleo-proteins, S1, S1-RBD, S2	Nasal swab, Nasopharyngeal swab	17 February 2021
			NL: Independent clinical field study, symptomatic individuals (n=417, PCR positive n=70), NP swab; sensitivity overall: 75.7%, sensitivity Ct _≤ 30: 85.2%, sensitivity Ct _≤ 25: 90.7%; specificity overall: 100%					
			NL: Independent clinical field study, symptomatic individuals (n=240, PCR positive n=21), NP+OP swab; sensitivity overall: 85.7%, sensitivity Ct _≤ 30: 89.5%, sensitivity Ct _≤ 25: 100%; specificity overall: 100%					
			NL: Independent clinical field study, symptomatic individuals (n=94, PCR positive n=18), NP+OP swab in VTM; sensitivity overall: 90.0%, sensitivity Ct _≤ 30: 100%, sensitivity Ct _≤ 25: 100%; specificity overall: 97.3%					
Hoyotek Biomedical Co.,Ltd.	Corona Virus (COVID-19) Antigen Rapid Test (Colloidal Gold)	1929	<i>Retrospective in vitro study</i>	NP swab - Sensitivity: 96%, Specificity: 99% OP swab - Sensitivity: 93%, Specificity: 97.5%	DE ^[2]	Unknown	Nasopharyngeal swab, Oropharyngeal swab	20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 90% at Ct _≤ 25; Manufacturer specificity: 99%					
Hubei Jinjian Biology Co., Ltd	SARS-CoV-2 Antigen Test Kit	1759	<i>Retrospective in vitro study</i>	Sensitivity: 98.02% Nasal Swab	DE ^[2]	Nucleo-protein	Nasopharyngeal swab	23 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct _≤ 25; Manufacturer specificity: 99.3%					
Humasis	Humasis COVID-19 Ag Test	1263	<i>Retrospective in vitro study</i>	95.3% sensitivity 100% specificity Nasal swab	DE ^[2] UK	Unknown	Nasal swab	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut (NP swab): Sensitivity of 88% at Ct _≤ 25; Manufacturer specificity: 100%					

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Innova Medical Group.Inc	Innova SARS-CoV-2 Antigen Rapid Qualitative Test	1801	<i>Retrospective in vitro study</i>	Sensitivity 94.0% : CI 95% (86.7%-98.0%) – calculated for viral loads $\times 10^6$ copies RNA /mL	DE ^[2]	Nucleo-capsid protein	Anterior nasal swab, Nasal swab	20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct \leq 25; Manufacturer specificity: 99%	Specificity: 99.6% - CI:95%(99.4%-99.8%)				
Innovation Biotech(Beijing) Co.Ltd	Coronavirus (SARS-Cov-2) Antigen Rapid Test Cassette (Nasal swab)	2278	<i>Retrospective in vitro study</i>	Sensitivity: 95.6% Specificity: 100%	DE ^[2]	Nucleo-protein	Nasal swab	20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct \leq 25; Manufacturer specificity: 99%					
InTec PRODUCTS, INC.	Rapid SARS-CoV-2 Antigen Test (nasopharyngeal specimen)	2419	<i>Retrospective in vitro study</i>	Sensitivity 90.2% (95% CI: 83.1% to 95.0%); Specificity 100.0% (95% CI: 96.5% - 100.00%)	DE ^[2]	Nucleo-capsid protein	Nasopharyngeal swab	20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct \leq 25; Manufacturer specificity: 100%					
Jiangsu Bioperfectus Technologies Co., Ltd.	Novel Corona Virus (SARS-CoV-2) Ag Rapid Test Kit	2107	<i>Retrospective in vitro study</i>	Sensitivity NP: 95.48% (95%CI:93.01%-96.01%), Nasal:95.33% (95%CI:91.31%-96.60%) Specificity NP: 99.61% (95%CI:97.85%-99.93%), Nasal:99.16% (95%CI:95.39%-99.85%)	DE ^[2] UK	Nucleo-protein	Nasal swab, Nasopharyngeal swab,	14 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct \leq 25; Manufacturer specificity: 99.15%					
Jiangsu Diagnostics Biotechnology Co., Ltd	COVID-19 Antigen Rapid Test Cassette (Colloidal Gold)	1920	<i>Retrospective in vitro study</i>	97.58 % sensitivity 100 % specificity Nasal/NP/ OP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab, Throat swab	14 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct \leq 25; Manufacturer specificity: 100%					
Jiangsu Medomics medical technology Co.,Ltd.	SARS-CoV-2 antigen Test Kit (LFIA)	2006	<i>Retrospective in vitro study</i>	Sensitivity: 97.73% Specificity: 99.51% Anterior nasal swab, NP swab	DE ^[2]	Nucleo-protein	Anterior nasal swab, Nasopharyngeal swab, Throat swab	7 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 94,1% at Ct \leq 25; Manufacturer specificity: 99,51%					
Jiangsu Mole Bioscience CO., LTD.	SARS-CoV-2 Antigen Test Cassette	2586	<i>Retrospective in vitro study</i>	sensitivity: 98.31 %, specificity: 99.17 %	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab	8 December 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct \leq 25; Manufacturer specificity: 99,17%					

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Jiangsu Well Biotech Co., Ltd.	COVID-19 Ag Rapid Test Device	2144	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99%	sensitivity: 94.74%, specificity: 99%	DE ^[2]	Nucleo-protein	Nasal swab	20 October 2021
Joinstar Biomedical Technology Co. Ltd	COVID-19 Rapid Antigen Test (Colloidal Gold)	1333	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.1%	96.1% sensitivity 98.1% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	17 February 2021
IEDAU INTERNATIONAL GMBH	Covid-19 Antigen Schnelltest (Colloidal Gold)	2555	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.2%	OP/Nasal: sensitivity: 96,1%, specificity: 99,2% NP: sensitivity: 97,1%, specificity: 99,2 %	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	8 December 2021
JOYSBIO (Tianjin) Biotechnology Co., Ltd.	SARS-CoV-2 Antigen Rapid Test Kit (Colloidal Gold immunochromatography)	1764	<i>Prospective clinical field studies</i> CZ N=225 (90 RT-PCR positive), 60.3% symptomatic patients. Test parameters for a subgroup of symptomatic patients (estimates and 95% confidence intervals): sensitivity 92% (80.8–97.8), specificity 97.6% (91.5–99.7). Test parameters for a subgroup of asymptomatic patients (estimates and 95% confidence intervals): sensitivity 100% 100 (54.1–100), specificity 100% (95.5–100). FIND Evaluation CH (11 Feb 2021) 265 samples, Nasal swab. Clinical sensitivities: Days ≤ 7: 74.2%; Ct ≤ 33: 78.9%; Ct ≤ 25: 91.3%; Clinical specificity: 99.1%	98.13% sensitivity Nasal swab	CZ, DE ^[2] CH	Unknown	Nasal swab	10 May 2021
Labnovation Technologies Inc.	SARS-CoV-2 Antigen Rapid Test Kit	1266	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 94% at Ct ≤ 25; Manufacturer specificity: 97.3%	96.3% sensitivity, 97.3% specificity NP/OP swab	DE ^[2]	Nucleo-protein	Nasopharyngeal swab, Oropharyngeal swab	10 May 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Lumigenex (Suzhou) Co., Ltd	PocRoc® SARS-CoV-2 Antigen Rapid Test Kit (Colloidal Gold)	2128	<i>Retrospective in vitro study</i>	93.33% sensitivity 99.16% specificity Nasal/NP/OP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99,16%					
LumiQuick Diagnostics Inc.	QuickProfile™ COVID-19 Antigen Test	1267	<i>Retrospective in vitro study</i>	93.7% sensitivity, 98.8% specificity NP swab	DE ^[2]	Unknown	Nasopharyngeal swab	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.8%					
LumiraDX	LumiraDx SARS-CoV-2 Ag Test	1268	<i>Prospective clinical field study</i>	97.6% sensitivity 96.6% specificity Nasal swab	DE ^[2] , ES SKUP CH	Nucleo-protein	Nasal swab	17 February 2021
			SKUP/2021/124: 448 samples: 83 positive samples and 365 negative samples. Nasal specimen: diagnostic sensitivity of 87% (79-92) and diagnostic specificity of 99,5% (98,3-99,9). NP specimen: diagnostic sensitivity of 90% (83-95) and diagnostic specificity of 97,8% (96,0-98,8) (Scandinavian evaluation of laboratory equipment for point of care testing)					
			FIND Evaluation DE (8 Oct 2021) 761 samples, NP swab. Clinical sensitivities: Days ≤ 7: 86.4%; Ct ≤ 33: 87.2%; Ct ≤ 25: 92.6%; Clinical specificity: 99.3%					
			Brazil (8 Oct 2021) 251 samples, NP swab. Clinical sensitivities: Days ≤ 7: 85.7%; Ct ≤ 33: 87.7%; Ct ≤ 25: 94.1%; Clinical specificity: 99%					
MEDsan GmbH	MEDsan SARS-CoV-2 Antigen Rapid Test	1180	<i>Retrospective in vitro study</i>	92.5% sensitivity 99.8% specificity NP/OP swab	DE ^[2] CH	Unknown	Nasopharyngeal swab, Oropharyngeal swab	17 February 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.8%					

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Merlin Biomedical (Xiamen) Co., Ltd.	SARS-CoV-2 Antigen Rapid Test Cassette	2029	<i>Retrospective in vitro study</i>	95.05% sensitivity 98.99% specificity Nasal/NP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab	16 June 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 90% at Ct ≤ 30 and 100% at Ct ≤ 25; Manufacturer specificity: 98.99%					
MEXACARE GmbH	MEXACARE COVID-19 Antigen Rapid Test	1775	<i>Retrospective in vitro study</i>	Sensitivity: 96.17% Specificity: 99,1% Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab	7 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99,1%					
möLab	mö-screen Corona Antigen Test	1190	<i>Retrospective in vitro study</i>	Sensitivity: 97.25% Specificity: 99.99% NP swab	DE ^[2] , IE	Unknown	Nasopharyngeal swab	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99,99%					
Mologic Ltd	COVIO Ag COVID-19 Antigen Rapid Diagnostic Test	2640	<i>Prospective clinical field study</i>	Sensitivity: 90.6%, Specificity: 100% Nasal swab	DE ^[2] UK	Nucleo-capsid protein	Nasal swab	8 December 2021
			FIND evaluation DE: Symptomatic and asymptomatic (n=649, PCR positive = 191), nasal and nasal-mouth-throat swab; sensitivity overall: 90.6%, sensitivity Ct ≤ 25: 96.4%; specificity: 100%					
MP Biomedicals	Rapid SARS-CoV-2 Antigen Test Card	1481	<i>Retrospective in vitro study</i>	96.17% sensitivity 99.16% specificity Nasal swab, Anterior nasal swab	DE ^[2] CH, UK	Nucleo-protein	Anterior nasal swab, Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	17 February 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.03%					
Nal von minden GmbH	NADAL COVID -19 Ag +Influenza A/B Test	2104	<i>Retrospective in vitro study</i>	97% sensitivity 98% specificity NP swab	DE ^[2]	Nucleo-protein	Nasopharyngeal swab	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 83% at Ct ≤ 25; Manufacturer specificity: 99.9%					
Nal von minden GmbH	NADAL COVID -19 Ag Test	1162	<i>Prospective clinical field study</i>	97.6% sensitivity 99.9% specificity Nasal swab	DE ^[2] , FR China	Nucleo-protein	Nasal swab † Serum, Whole blood	17 February 2021
			FIND evaluation CH (26 April 2021) 462 samples, NP swab. Clinical sensitivities: Days ≤ 7: 88.5%; Ct ≤ 33: 92.4%; Ct ≤ 25: 97.8%; Clinical specificity: 99.2%					

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
			<i>Retrospective in vitro study</i>					
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 83% at Ct ≤ 25; Manufacturer specificity: 99.9%					
Nanjing Liming Bio-Products Co., Ltd.	StrongStep® SARS-CoV-2 Antigen Rapid Test	2301	<i>Retrospective in vitro study</i>	Sensitivity: 96.19 %, Specificity: 99.26 % Nasal swab	DE ^[2]	Nucleo-capsid protein	Nasal swab	8 December 2021
			<i>Retrospective in vitro study</i>					
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.26%					
Nanjing Norman Biological Technology Co., Ltd.	Novel Coronavirus (2019-nCoV) Antigen Testing Kit (Colloidal Gold)	2506	<i>Retrospective in vitro study</i>	Clinical sensitivity: - 91.13 % (Saliva) - 93.02 % (Anterior Nasal swab) - 93.21 % (NP swab) Clinical specificity: - 93.02 % (Anterior Nasal swab) - 99.23 % (Anterior Nasal swab) - 99.29 % (NP swab)	DE ^[2]	Nucleo-capsid protein	Anterior nasal swab, Nasopharyngeal swab, Oropharyngeal swab ! Saliva	10 November 2021
			<i>Retrospective in vitro study</i>					
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 88% at Ct ≤ 25; Manufacturer specificity: 100%					
NanoEntek	FREND COVID-19 Ag	1420	<i>Retrospective in vitro study</i>	94.12% sensitivity 100% specificity NP swab	DE ^[2]	Nucleo-protein	Nasopharyngeal swab	10 May 2021
			<i>Retrospective in vitro study</i>					
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 94,1% at at Ct ≤ 25; Manufacturer specificity: 98.4%					
NanoRepro AG	NanoRepro SARS-CoV-2 Antigen Rapid Test	2200	<i>Retrospective in vitro study</i>	97.2 % sensitivity 98.4% specificity Nasal/NP/OP swab	DE ^[2]	Nucleo-protein	Anterior nasal swab, Nasopharyngeal swab, Oropharyngeal swab	14 July 2021
			<i>Prospective clinical field study</i>					
			ES: Independent validation study; Nasal test compared to nasal PCR. Sensitivity 95.24% (Ct<30), Specificity 100%.					
NESAPOR EUROPA SL	MARESKIT COVID-19 ANTIGEN RAPID TEST KIT	2241	<i>Prospective clinical field study</i>	Sensitivity: 95.24% (95% CI: 83.84% to 99.42%), Specificity: 100% (95% CI: 97.22% to 100.00%) Nasal swab	ES	Nucleo-protein	Nasal swab	23 July 2021
			<i>Retrospective in vitro study</i>					
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 92,5% at Ct ≤ 30 and 100% at Ct ≤ 25; Manufacturer specificity: 99.2%					
New Gene (Hangzhou) Bioengineering Co., Ltd.	COVID-19 Antigen Detection Kit	1501	<i>Retrospective in vitro study</i>	98% sensitivity 99.2% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab ! Saliva, Sputum	16 June 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
NG Biotech	Ninonasal	1880	<i>Prospective clinical field study</i>	Clinical sensitivity: 98%, Clinical specificity: 99%	FR	Nucleo-protein	Nasal swab, Nasopharyngeal swab	10 November 2021
			FR: Prospective validation study for NP and nasal swabs: NP sensitivity 89% (75/84) , specificity 99% (92/93). Nasal sensitivity 98% (125/128), specificity 99% (388/390)					
Novatech	SARS-CoV-2 Antigen Rapid Test	1762	<i>Retrospective in vitro study</i>	95 % sensitivity 100% specificity Nasal/ NP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab	14 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 100%					
Oncosem Onkolojik Sistemler San. ve Tic. A.S.	CAT	1199	<i>Retrospective in vitro study</i>	93.75% sensitivity 98.04% specificity Nasal swab	DE ^[2]	Unknown	Nasal swab	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 98,04%					
PCL Inc.	PCL COVID19 Ag Rapid FIA	308	<i>Prospective clinical field study</i>	94,92% sensitivity, 99,99% specificity	DE ^[2] , FR	Unknown	Nasopharyngeal Swab	10 May 2021
			FR: Validation study data: NP swabs, sensitivity 94.29% (33/35) and specificity 100% (70/70)					
PCL Inc.	PCL COVID19 Ag Gold	2243	<i>Prospective clinical field study</i>		FR	Nucleo-protein	Nasal swab, Nasopharyngeal swab ! Saliva	7 July 2021
			FR: Validation study data: 120 positive and 200 negative samples; sensitivity 92%, specificity: 100%					
PerGrande Bio Tech Development Co., Ltd.	SARS-CoV-2 Antigen Detection Kit (Colloidal Gold Immunochromatographic Assay)	2116	<i>Retrospective in vitro study</i>	94.28% sensitivity 99.11% specificity NP/Nasal/OP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.11%					
Precision Biosensor Inc.	Exdia COVI-19 Ag	1271	<i>Retrospective in vitro study</i>	93.9% sensitivity 98% specificity NP swab	DE ^[2] CH	Unknown	Nasopharyngeal swab	17 February 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.3%					

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Prognosis Biotech	Rapid Test Ag 2019-nCov	1495	<i>Retrospective in vitro study</i>	Clinical Sensitivity: 95.56 % Specificity: 99,58% Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab	7 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 99,58%					
Qingdao Hightop Biotech Co. Ltd	SARS-CoV-2 Antigen Rapid Test (Immunochromatography)	1341	<i>Retrospective in vitro study</i>	95% sensitivity 99.75% specificity Nasal swab	DE ^[2]	Nucleo-protein	Anterior nasal swab, Nasal swab	17 February 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 30 and 100% at Ct ≤ 25; Manufacturer specificity: 99.75%					
Quidel Corporation	Sofia SARS Antigen FIA	1097	<i>Prospective clinical field studies</i>	96.7% sensitivity 100% specificity NP/Nasal swab	DE ^[2] , NL ^[5] , PT CH	Nucleo-protein	Nasal swab, Nasopharyngeal swab	17 February 2021
			FR: Validation study data: NP swabs sensitivity 84,44% (76/90), specificity 99,19 (491/495)					
			NL: Independent prospective clinical field study in symptomatic (n=733, PCR positive 144); NP swab; sensitivity overall: 84.0%, sensitivity Ct<30: 90.1%, sensitivity Ct<25: 92.5%; specificity overall: 99.8%.					
			PT: 80 samples from symptomatic individuals (27 PCR positive and 53 negative by PCR) were tested. Sensitivity 70% (95%IC50-86); specificity 100% (95%IC 93-100). TCID50/ml 0,68x 102 and CT<25.					
			<i>Retrospective in vitro study</i>					
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 89% at Ct ≤ 25; Manufacturer specificity: 100%					
Rapid Pathogen Screening, Inc	LIAISON® Quick Detect Covid Ag Assay	2290	<i>Retrospective in vitro study</i>	Sensitivity: 96.1%, Specificity: 97% NP and Nasal swab	IT	Nucleo-protein	Nasal swab, Nasopharyngeal swab	23 July 2021
			IT: Independent validation study, 100 pos. and 100 neg. samples; sensitivity: 92.7% with Ct<25; specificity: 100%.					

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS- CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Roche (SD BIOSENSOR)	SARS-CoV-2 Rapid Antigen Test	1604	<i>Prospective clinical field study</i>	96.52% sensitivity 99.2% specificity NP swab	DE ^[2] , FI, NL, PT, SE UK	Nucleo- protein	Nasopharyngeal swab	10 May 2021
			NL: Independent prospective clinical field study in symptomatic (n=970, PCR positive 186); NP swab; sensitivity overall: 84.9%, sensitivity Ct≤30: 94.3%, sensitivity Ct≤25: 99.1%; specificity overall: 99.5%					
			SE: Karolinska hospital evaluation of Lot QCO3020109. Patient samples: 95 PCR positive, 150 negative. No detailed sample description available. Sensitivity 43%, specificity 100%. Sensitivity Ct<25 = 80.5%.					
Roche (SD BIOSENSOR)	SARS-CoV-2 Rapid Antigen Test Nasal	2228	<i>Retrospective in vitro studies</i>	Clinical Sensitivity: 89.6 % (Ct ≤ 30) 93.1 % (Ct ≤ 27) Clinical Specificity: 99.1 % Nasal swab	DE ^[2] Brazil, UK	Nucleo- protein	Nasal swab	7 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 89% at Ct< 25; Manufacturer specificity: 99.68%					
			FIND evaluation DE (12 April 2021) 179 samples, nasal swab. Clinical sensitivities: Days ≤7: 81.2%; Ct ≤33: 87.5%; Ct ≤ 25: 100%; Clinical specificity: 99.3% Brazil (12 April 2021) 214 samples, nasal swab. Clinical sensitivities: Days ≤7: 81.2%; Ct ≤33: 91.7%; Ct ≤ 25: 100%; Clinical specificity: 99.3%					
Safecare Biotech (Hangzhou) Co. Ltd	COVID-19 Antigen Rapid Test Kit (Swab)	1489	<i>Retrospective in vitro study</i>	97.27% sensitivity 99.42% specificity Nasal swab	DE ^[2]	Nucleo- protein	Nasal swab	17 February 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.42%					

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
	Multi-Respiratory Virus Antigen Test Kit (Swab) (Influenza A+B/COVID-19)	1490	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct < 25; Manufacturer specificity: 99.44%	97.04% sensitivity 99.44% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab	10 May 2021
ScheBo Biotech AG	ScheBo SARS CoV-2 Quick Antigen	1201	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 95% at Ct < 30 and 100% at Ct < 25; Manufacturer specificity: 99%	96.6% sensitivity (Ct ≤ 30) 99.00% specificity NP/ OP swab	DE ^[2]	Nucleo-protein	Nasopharyngeal swab, Oropharyngeal swab ! Serum	16 June 2021
SD Biosensor Inc	STANDARD Q COVID-19 Ag Test Nasal	2052	<i>Prospective clinical field studies</i> DE 146 symptomatic adults, 40 (27.4%) were RT-PCR-positive for SARS-CoV-2. Sensitivity with 85.0% (34/40; 95% CI 70.9-92.9) with professional testing. At high viral load (>7.0 log10 SARS-CoV-2 RNA copies/ml), sensitivity was 96.6% (28/29; 95% CI 82.8-99.8) for professional testing. FIND evaluation DE (12 April 2021) 179 samples, nasal swab. Clinical sensitivities: Days ≤ 7: 81.2%; Ct ≤ 33: 87.5%; Ct < 25: 100%; Clinical specificity: 99.3% Brazil (12 April 2021) 214 samples, nasal swab. Clinical sensitivities: Days ≤ 7: 81.2%; Ct ≤ 33: 91.7%; Ct < 25: 100%; Clinical specificity: 99.3%	Clinical Sensitivity: 97.12 % Clinical Specificity: 100 % Nasal swab	DE ^[2] , FI, FR Brazil, UK	Nucleo-protein	Nasal swab	7 July 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
SD BIOSENSOR Inc.	STANDARD F COVID-19 Ag FIA	344	<i>Prospective clinical field studies</i>	94,09% sensitivity 98.52% specificity NP swab	DE ^[2] , IT, NL ^[5] , DK Brazil, CH, India, UK	Nucleo-protein	Nasopharyngeal swab	17 February 2021
			NL Independent prospective clinical field study in symptomatic (n=628, PCR positive 118); NP swab; sensitivity overall: 78.0%, sensitivity Ct<30: 84.4%, sensitivity Ct<25: 90.3%; specificity overall: 99.6%					
			FIND evaluation DE (10 Dec 2020) 676 samples, NP swab. Clinical sensitivities: Days ≤ 7: 81.2%; - Ct ≤ 33: 75%; Ct ≤ 25: 100%; Clinical specificity: 96.9% Brazil (10 Dec 2020) 453 samples, NP swab. Clinical sensitivities: Days ≤ 7: 80.2%; - Ct ≤ 33: 80.9%; Ct ≤ 25: 87.9%; Clinical specificity: 97.9% India (25 June 2020) 417 samples, NP swab. Clinical sensitivities: Days ≤ 7: 61.8%; - Ct ≤ 33: 53.6%; Ct ≤ 25: 68.5%; Clinical specificity: 99.5%					
			<i>Retrospective in vitro study</i>					
SD BIOSENSOR Inc.	STANDARD Q COVID-19 Ag Test	345	DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.52%	96.52% sensitivity 99.68% specificity NP swab	DE ^[2] , ES, IT, NL ^[5] , DK, PT Brazil, CH, India, NO, UA, UK	Unknown	Nasopharyngeal swab	17 February 2021
			<i>Prospective clinical field studies</i> PT 80 samples from symptomatic individuals (27 PCR positive and 53 negative by PCR) were tested. Sensitivity 70% (95%IC50-86); specificity 100% (95%IC 93-100). TCID50/ml 0,68x 10 ² and CT<25. FIND evaluation DE (10 Dec 2020) 1263 samples, NP swab. Clinical sensitivities: Days ≤ 7: 80%; - Ct ≤ 33: 87.8%; Ct ≤ 25: 100%; Clinical specificity: 99.3% Brazil (10 Dec 2020)					

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
			<p>400 samples, NP swab. Clinical sensitivities: Days \leq 7: 90.7%; - Ct \leq 33: 91.9%; Ct \leq 25: 95.9%; Clinical specificity: 97.6% CH (10 Dec 2020)</p> <p>529 samples, NP swab. Clinical sensitivities: Days \leq 7: 89.8%; - Ct \leq 33: 91.8%; Ct \leq 25: 97.2%; Clinical specificity: 99.7% India (22 April 2021)</p> <p>334 samples, NP swab. Clinical sensitivities: Days \leq 7: 58.3%; - Ct \leq 33: 65.5%; Ct \leq 25: 89.4%; Clinical specificity: 97.3% Peru (22 April 2021)</p> <p>335 samples, NP swab. Clinical sensitivities: Days \leq 7: 81.4%; - Ct \leq 33: 83.3%; Ct \leq 25: 96.2%; Clinical specificity: 99.6%</p> <p><i>Retrospective in vitro studies</i></p> <p>DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 89% at Ct \leq 25; Manufacturer specificity: 99.68%</p>					
SGA Medikal	V-Chek SARS-CoV-2 Ag Rapid Test Kit (Colloidal Gold)	1319	<p><i>Retrospective in vitro study</i></p> <p>DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 94,1% at Ct \leq 25; Manufacturer specificity: 99,5%</p>	96.6% sensitivity, 99.5% specificity, Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab	10 May 2021
	V-Chek SARS-CoV-2 Rapid Ag Test (colloidal gold)	1357	<p><i>Retrospective in vitro study</i></p> <p>DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 94,1% at Ct \leq 25; Manufacturer specificity: 99,5%</p>	96.60% sensitivity: 99.5% specificity, Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab	7 July 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Shenzhen Ultra-Diagnostics Biotec Co., Ltd	SARS-CoV-2 Antigen Test Kit	2017	<i>Prospective clinical field study</i>	Clinical Sensitivity: 95.33 % (Nasal), 95.48(NP) Clinical Specificity: 99.16 % (Nasal), 99.61 % (NP)	BE, SI	Nucleo-protein	Nasal swab, Nasopharyngeal swab ! Saliva	10 May 2021
			SI: Sensitivity in unselected symptomatic population: 86.4% (172 RAT pos. / 199 RT-PCR pos.), sensitivity of 97.8% at Ct≤25. Specificity: 99.1% (1972 RAT neg. / 1990 RT-PCR neg.), NP swab					
Shenzhen CAS-Envision Medical Technology Co., Ltd.	SARS-CoV-2-Antigen Rapid Detection Kit	2152	<i>Retrospective in vitro study</i>	OP: Sensitivity: 98.1% 94.7%-99.4%), Specificity: 99.5% 97.0%-99.9%) NP: Sensitivity: 98.1% 94.7%-99.4%), Specificity: 99.5% 97.0%-99.9%)	DE ^[2]	Nucleo-capsid protein	Oropharyngeal swab; Nasopharyngeal swab	8 December 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct < 25; Manufacturer specificity: 99.5%					
Shenzhen Dymind Biotechnology Co., Ltd	SARS-CoV-2 Antigen Test Kit (Colloidal Gold)	2415	<i>Retrospective in vitro study</i>	Sensitivity: 96.58%, Specificity: 98.37%	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab	20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct < 25; Manufacturer specificity: 96.58%					
Shenzhen Huian Biosci Technology Co., Ltd.	SARS-CoV-2 Antigen Test Kit (Colloidal Gold)	2414	<i>Retrospective in vitro study</i>	NP/OP swab: Sensitivity: 95.0%, Specificity: 99.1% Nasal swab: Sensitivity: 94.6%, Specificity: 99.1%	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab	20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.1%					
Shenzhen Kisshealth Biotechnology Co., Ltd	SARS-CoV-2 Antigen Test Kit (GICA)	1813	<i>Retrospective in vitro study</i>	NP swabs: Sensitivity: 96.43%, Specificity: 100%. Nasal (Anterior) swabs: Sensitivity: 99.43%, Specificity: 99.23%.	DE ^[2]	Nucleo-capsid protein	Anterior nasal swab, Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.2%					
Shenzhen Lvshiyuan Biotechnology Co., Ltd.	Green Spring SARS-CoV-2 Antigen-Rapid test-Set	2109	<i>Retrospective in vitro study</i>	96.43% sensitivity 100% specificity NP/OP/Nasal swab	DE ^[2]	Nucleo-protein	Anterior nasal swab, Nasal swab, Nasopharyngeal swab, Oropharyngeal swab ! Saliva	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%					

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Shenzhen Microprofit Biotech Co., Ltd	SARS-CoV-2 Antigen Test Kit (Colloidal Gold Chromatographic Immunoassay)	1967	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%	Sensitivity: 92.93% Clinical Specificity: 100 % Nasal/NP/OP swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	7 July 2021
Shenzhen Microprofit Biotech Co., Ltd.	SARS-CoV-2 Spike Protein Test Kit (Colloidal Gold Chromatographic Immunoassay)	1178	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct < 25; Manufacturer specificity: 100%	Sensitivity: 86.3%, Specificity: 100% Nasal Swab	DE ^[2]	Spike protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	23 July 2021
Shenzhen Microprofit Biotech Co., Ltd	SARS-CoV-2 Spike Protein Test Kit (Fluorescence Immunoassay)	1228	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct < 25; Manufacturer specificity: 100%	Sensitivity: 93.46%, Specificity: 100%	DE ^[2]	Nucleo-protein, S protein (S1)	Nasopharyngeal swab	8 December 2021
Shenzhen Reagent Technology Co.,Ltd.	SARS-CoV-2 antigen IVD kit SWAB	2026	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct < 25; Manufacturer specificity: 98.1%	Sensitivity: 95.2 %, specificity: 98.1 %	DE ^[2]	Nucleo-protein	Nasopharyngeal swab, Oropharyngeal swab	20 October 2021
Shenzhen Watmind Medical Co., Ltd	SARS-CoV-2 Ag Diagnostic Test Kit (Colloidal Gold)	1769	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct < 25; Manufacturer specificity: 99.12%	Sensitivity: 95.15% (for symptom onset within 7 days) Specificity: 99.12% Nasal swab	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	10 May 2021
Shenzhen Watmind Medical Co., Ltd	SARS-CoV-2 Ag Diagnostic Test Kit (Immuno-fluorescence)	1768	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99,13%	Clinical Sensitivity: 97.83 % (CT ≤ 33) Clinical Sensitivity: 90.08 % (Ct ≤ 36) Specificity: 99,13% Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab	7 July 2021
Shenzhen YHLO Biotech Co., Ltd.	GLINE-2019-nCoV Ag	1347	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 95% at Ct ≤ 25; Manufacturer specificity: 99.85%	Nasal: Sensitivity: 97.37% (95%CI: 92.50% - 99.45%) Specificity: 99.25% (95%CI: 97.82% - 99.85%) NP: Sensitivity: 96.49% (95%CI: 91.26% - 99.04%) Specificity: 99.25% (95%CI: 97.82% - 99.85%)	DE ^[2]	Nucleo-capsid protein	Nasal swab; Nasopharyngeal swab	8 December 2021

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Shenzhen Zhenrui Biotech Co., Ltd	Zhenrui ®COVID-19 Antigen Test Cassette	1574	<i>Retrospective in vitro study</i>	96% sensitivity 97% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab ! Saliva	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 82% at Ct ≤ 25; Manufacturer specificity: 97%					
Sugentech, Inc.	SGTi-flex COVID-19 Ag	1114	<i>Retrospective in vitro study</i>	100% sensitivity 100% specificity OP/NP swab	DE ^[2]	Unknown	Nasopharyngeal swab, Oropharyngeal swab	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 30 and 100% at Ct ≤ 25; Manufacturer specificity: 99.0%					
SureScreen Diagnostics	SARS-CoV-2 Rapid Antigen Test Cassette	2297	<i>Retrospective in vitro study</i>	Sensitivity: 96.1%, Specificity: 99%	DE ^[2] UK	Nucleo-protein	Anterior nasal swab ! Other	20 October 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99%					
TODA PHARMA	TODA CORONADIAG Ag	1466	<i>Prospective clinical field study</i>	98.6% sensitivity Nasal swab	DE ^[2] , FR	Nucleo-protein	Nasal swab, Nasopharyngeal swab	10 May 2021
			FR: Validation data: NP swabs, sensitivity : 96,1-100%, specificity 99,2-100%					
			<i>Retrospective in vitro study</i>					
Triplex International Biosciences Co., Ltd	SARS-CoV-2 Antigen Rapid Test Kit	2074	<i>Retrospective in vitro study</i>	98.51% sensitivity 99.91% specificity Nasal/OP/NP swab	DE ^[2]	Nucleo-capsid protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab ! Saliva	16 June 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 92,5% at Ct ≤ 30 and 100% at Ct ≤ 25; Manufacturer specificity: 99.91%					
Triplex International Biosciences Co., Ltd, China	SARS-CoV-2 Antigen Rapid Test Kit	1465	<i>Retrospective in vitro study</i>	98.51 % sensitivity 100% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab	14 July 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%					

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Vitrosens Biotechnology Co., Ltd	RapidFor SARS-CoV-2 Rapid Ag Test	1443	<i>Retrospective in vitro study</i>					
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 30 and 100% at Ct ≤ 25; Manufacturer specificity: 99.05%	97.30% sensitivity 99.05% specificity NP swab	DE ^[2]	Nucleo-protein	Anterior nasal swab, Nasal swab, Nasopharyngeal swab, Oropharyngeal swab, Throat swab	10 May 2021
VivaChek Biotech (Hangzhou) Co., Ltd.	VivaDiag Pro SARS-CoV-2 Ag Rapid Test	2103	AT: 97,06% sensitivity, 100% specificity, all specimen types, i.e. N&OP&NP swab	Sensitivity: nasal: 90.38%, oropharyngeal: 91.11%, nasopharyngeal: 90.6% Specificity: nasal: 99.99%, oropharyngeal: 99.99%, nasopharyngeal: 99.99%,	AT, DE ^[2] , SI	Unknown	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	10 May 2021 ²²
Wuhan EasyDiagnosis Biomedicine Co., Ltd.	COVID-19 (SARS-CoV-2) Antigen-Test Kit	2098	<i>Retrospective in vitro study</i>					
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.26%	96.1% sensitivity 100% specificity Nasal/OP/NP swab	DE ^[2] UK	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	10 May 2021
Wuhan Life Origin Biotech Joint Stock Co., Ltd.	SARS-CoV-2 Antigen Assay Kit (Immuno-chromatography)	1773	<i>Retrospective in vitro study</i>					
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: %	92.67% sensitivity Nasal swab	DE ^[2]	Unknown	Nasal swab	14 July 2021
Wuhan UNscience Biotechnology Co., Ltd.	SARS-CoV-2 Antigen Rapid Test Kit	2090	<i>Retrospective in vitro study</i>					
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99,57%	Sensitivity: 96.33% Specificity: 99.57% Nasal/NP/OP swab	DE ^[2] , FR	Nucleo-protein	Mid-turbinate swab, Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	7 July 2021
Wuxi Biohermes Bio & Medical Technology Co., Ltd.	SARS-CoV-2 Antigen Test Kit (Lateral Flow Assay)	2143	<i>Retrospective in vitro study</i>					
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.02%	Sensitivity: 95.31 %, Specificity: 98.02 %	DE ^[2]	Nucleo-protein	Nasopharyngeal swab, Oropharyngeal swab	20 October 2021

²² This rapid antigen test, device ID 2103, was removed from the EU common list on 10 November 2021. The grace period will end on 5 January 2022, 23:59 CET

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Xiamen AmonMed Biotechnology Co., Ltd	COVID-19 Antigen Rapid Test Kit (Colloidal Gold)	1763	Retrospective in vitro study	93.2% sensitivity 99.55% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab	10 May 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.55%					
Xiamen Boson Biotech Co. Ltd	Rapid SARS-CoV-2 Antigen Test Card	1278	Retrospective in vitro study	96.49% sensitivity 99.03% specificity NP swab	DE ^[2] CH, UK	Unknown	Nasopharyngeal swab	17 February 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.03%					
Xiamen Wiz Biotech Co., Ltd	SARS-CoV-2 Antigen Rapid Test	1456	Retrospective in vitro study	96.3% sensitivity, 100% specificity Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab	10 May 2021
	SARS-CoV-2 Antigen Rapid Test (Colloidal Gold)	1884	Retrospective in vitro study				DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 88% at Ct ≤ 25; Manufacturer specificity: 100%	
Zhejiang Anji Saianfu Biotech Co., Ltd	AndLucky COVID-19 Antigen Rapid Test	1296	Retrospective in vitro study	95.8% sensitivity, 99% specificity, Nasal swab	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab, Oropharyngeal swab	10 May 2021
Zhejiang Anji Saianfu Biotech Co., Ltd	reOpenTest COVID-19 Antigen Rapid Test	1295	Retrospective in vitro study				DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 94,1% at Ct ≤ 25; Manufacturer specificity: 99%	
Pantest SA	Pantest Coronavirus Ag (Nasopharyngeal Swab)	2271	Retrospective in vitro study	sensitivity: 95,70%, specificity: 99,10%	DE ^[2]	Nucleo-capsid protein	Nasopharyngeal swab	8 December 2021
			DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 94% at Ct ≤ 25; Manufacturer specificity: 99.1%					

Manufacturer	RAT commercial name	Device ID # ¹⁵	Clinical performance <i>As reported by independent validation studies</i>	Clinical performance <i>Data by manufacturer</i> ¹⁶	Completed validation studies	SARS-CoV-2 Target protein	Specimen ¹⁷	Included in EU common list since:
Zhejiang GENE SCIENCE Co., Ltd	Novel Coronavirus (COVID-19) Antigen Detection Kit (Swab)	2684	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 98.73%	OP: Sensitivity: 95.65%, Specificity: 99.17% NP: Sensitivity: 94.58%, Specificity: 98.73%	DE ^[2]	Nucleo-capsid protein	Oropharyngeal swab; Nasopharyngeal swab	8 December 2021
Zhejiang Orient Gene Biotech Co., Ltd	Coronavirus Ag Rapid Test Cassette (Swab)	1343	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 99.22%	98.32 % sensitivity 99.6 % specificity Nasal/NP swab	DE ^[2] UK	Nucleo-protein	Nasal swab, Nasopharyngeal swab	17 February 2021
Zhuhai Encode Medical Engineering Co.,Ltd	ENCODE SARS-COV-2 Antigen Rapid Test Device	1902	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 95% at Ct ≤ 25; Manufacturer specificity: 100%	Throat swab/Nasal Swab: Sensitivity 96.49%, Specificity 100% Anterior Swab: Sensitivity 94.74%, Specificity: 100%	DE ^[2] UK	Nucleo-capsid protein	Anterior nasal swab, Nasal swab, Throat swab	20 October 2021
Zhuhai Lituo Biotechnology Co., Ltd.	COVID-19 Antigen Detection Kit (Colloidal Gold)	1957	<i>Retrospective in vitro study</i> DE: Positive evaluation by Paul-Ehrlich-Institut: Sensitivity of 100% at Ct ≤ 25; Manufacturer specificity: 100%	96.12% sensitivity Nasal swab (CT≤33); 99.59% sensitivity NP swab; 100% specificity Nasal swab (CT≤33)	DE ^[2]	Nucleo-protein	Nasal swab, Nasopharyngeal swab	14 July 2021

Notes:

[1] FR: Reference to validation study (not specifying which specific RAT is being recommended or was tested in practice): https://www.has-sante.fr/upload/docs/application/pdf/2020-10/synthese_tests_antigeniques_vd.pdf

[2] DE: Rapid antigen tests that have completed practical validation studies in Germany: See: <https://www.pei.de/SharedDocs/Downloads/DE/newsroom/dossiers/evaluierung-sensitivitaet-sars-cov-2-antigentests-04-12-2020.pdf?blob=publicationFile&v=43>

[3] SE: Smaller evaluations ongoing in some of the regions.

[4] BE: In the clinical performance study performed in three different clinical laboratories during the ascendant phase of the epidemiological curve, we found an overall sensitivity and specificity of 57.6 and 99.5%, respectively with an accuracy of 82.6%.

[5] NL: Collected validation data from accredited laboratories in the Netherlands. The report includes evaluations of various RAT that labs performed at their own initiative.

<https://lci.rivm.nl/antigeensneltesten>

[6] BE: Van Honacker E. et al., Comparison of five SARS-CoV-2 rapid antigen detection tests in a hospital setting and performance of one antigen assay in routine practice: a useful tool to guide isolation precautions? J Hosp Infect. In press.